



# SKC

SKC Inc., 863 Valley View Road,  
Eighty Four, PA 15330 USA

Phone 724-941-9704  
Fax 724-941-1369  
www.skcinc.com

## SKC AirChek 2000 CE Conformity Declaration

**Date:** October 2010

**In Accordance with:** Directive 89/336/EEC

SKC AirChek 2000 sample pumps conform to CE requirements and test to the EMI (electromagnetic interference) standard because they fall below the CE Low Voltage Standard (voltage rating between 50 and 1,000 volts for alternating current and between 75 and 1,500 volts for direct current).

SKC AirChek 2000 sample pumps are in accordance with Directive 89/336/EEC, and conform to the following standards:

EN 55011  
EN 61000-4-3  
ENV 50204  
EN 61000-4-2

Signed:

A handwritten signature in black ink that reads "Scott Marshall". The signature is written in a cursive, flowing style.

Scott Marshall, ASQ-CMQ/OE  
Quality Administrator  
SKC Inc.