



Bios

Driving a Higher Standard
in Flow MeasurementSM

Bios International CE Conformity Declaration

Effective January 1, 2007

In accordance with EMC (Electro Magnetic Compatibility) Directive 89/336/EEC

Bios products conform to CE requirements and test to the EMI standard (electromagnetic interference) because they fall below the CE Low Voltage standard (voltage rating between 50 and 1,000V for alternating current and between 75 and 1,500V for direct current).

In accordance with EMC Directive 89/336/EEC, Bios products conform to the following standards:

NF EN 61000-4-2-CEI 801-2

NF EN 61000-4-4-CEI 801-4

NF EN 55022

NF C 46-022-CEI 801-3