




863 Valley View Rd, Eighty Four, PA 15330
724-941-9701 • skcinc.com

Impinger Holders for AirChek Touch Series Sample Pumps Cat. No. 225-20 Series Operating Instructions

Attach Glass Midget Impinger Holder Cat. No. 225-20-03 or 225-20-05 or PFA Midget Impinger Holder Cat. No. 225-20-04 or 225-20-06 to the pump belt clip as follows. **Note:** Cat. No. 225-20-3 is shown below.

1. Turn the holder on its side and insert the toggle on the back into the slot in the pump belt clip.



2. Rotate the holder 90 degrees so that the toggle catches on either side of the slot and is secure. The bottom plate of the holder should be parallel with the bottom of the pump.



SKC Limited Warranty and Return Policy

SKC products are subject to the SKC Limited Warranty and Return Policy, which provides SKC's sole liability and the buyer's exclusive remedy. To view the complete SKC Limited Warranty and Return Policy, go to skcinc.com/warranty.