



Operating Instructions

863 Valley View Road, Eighty Four, PA 15330
724-941-9701 • skcinc.com

SKC User-loaded Disposable Parallel Particle Impactors (PPI)

The patented[†] impaction-based SKC Disposable Parallel Particle Impactor (PPI[®]) Samplers are designed to match precisely the collection efficiency curves for respirable and thoracic dust specified by ISO 7708/CEN and adopted by ACGIH, CEN, and other occupational hygiene organizations. The PPI Samplers are listed in the OSHA and MSHA final silica rules as meeting ISO 7708 criteria. The thoracic model meets the requirements of compounds with ACGIH[®] thoracic TLV[®]s.



Disposable PPI Samplers offer the convenience of single use to eliminate sampler assembly and cleaning; small size for worker comfort, even under helmets or other PPE; and a choice of flow rates for maximum flexibility in pump options, sample duration, and contaminant concentration. **These instructions are for PPI models that are NOT preloaded with filters and supports; these models contain only impaction substrates.**

Performance Profile

Sampling Rate: 2 L/min respirable or thoracic and 4 or 8 L/min respirable

Sample Pump: • Universal or AirChek[®] Series for 2 and 4 L/min
• Leland Legacy for 8 L/min

Sample Time: *Dependent on method used. Note: SKC tests indicate that a particulate mass of up to 6.8 mg on the four impaction substrates would not affect PPI performance. This amount is equivalent to sampling for 6 hours at 4 L/min in environments where respirable mass concentration is 4.76 mg/m³ and equals 50% of total dust. However, labs have reported to SKC that they prefer no more than 2 mg on the filter for analytical reasons. Therefore, SKC recommends that you work with your lab to determine optimum sample times for your unique sampling conditions.*

Sample Media (recommended): 37-mm, 5.0- μ m PVC filter or 37-mm, 0.8- μ m MCE filter

Impaction Substrate: Use cellulose or stainless steel screen for support
Four 3/8-in diameter pre-oiled porous plastic discs
(assembled in all Disposable PPIs)

Analysis: Gravimetric and/or chemical

Body Material: Conductive ABS plastic

Dimensions:

Height: 4.25 in (10.8 cm) - clip to exhaust

Diameter: 1.8 in (4.6 cm)

Depth: 1.2 in (3 cm)


Weight: 1.1 oz (31.2 gm)

Shelf-life: Limited; check expiration date on packaging


[†] U.S. Patent No. 7,073,402

Principle of Operation


SKC Disposable PPI Samplers are impaction-based filter samplers that perform precise size-selection for either thoracic or respirable dust, depending on the model. PPI Samplers contain four small impactors in the inlet section of the device. Each impactor features a unique 50% cut-point to target a specific one-quarter segment of the ISO/CEN curve resulting in a precise fit along the entire curve. A sample pump operating at 2, 4, or 8 L/min (2 L/min only for thoracic) pulls air through the inlet nozzle of each impactor in the inlet plate. Particles larger than each impactor's 50% cut-point are scrubbed and retained by impaction onto the porous oiled impaction substrate contained in each impactor. Smaller particles continue to the standard 37-mm collection filter for analysis.



PPI Performance



Collection efficiency of the 2, 4, and 8 L/min respirable PPI Samplers compared to the ISO respirable curve



Collection efficiency of the 2 L/min thoracic PPI Sampler compared to the ISO thoracic convention

Media and Sampler Preparation

Media Preparation

Condition and weigh filters according to method used. Record the weight as the pre-sample weight.

Sampler Preparation

The PPI will arrive with the inlet plate shrink wrapped and base plate separated.

1. Remove shrink wrap from inlet plate.



2. Using forceps, insert first a 37-mm support pad and then a conditioned and weighed 37-mm collection filter (if using gravimetric analysis) into the base plate.



3. Lay base plate on a flat surface, align inlet plate with base, and press down firmly to achieve an even seal.



4. Write sample ID on sampling label. Adhere sampling label to **bottom** of base plate.



5. Remove protective end cap from exhaust.

Technical Tidbits:

- Use forceps to carefully insert or remove collection filter and support pad. See *Accessories* for forceps.

Flow Rate Verification and Sampling

Tip As the particle load on the filter increases during sampling, the pressure drop will also increase. Therefore, use a compensating sample pump such as the AirChek Series or Leland Legacy depending on flow rate requirements.


Flow Rate Verification

Verify pump flow rate with a representative Disposable PPI (support and filter loaded) in line. **Note:** If using SKC High Flow chek-mate Flowmeter, Pulsation Dampener Cat. No. 375-150 is also required in line. **See pump and flowmeter operating instructions.**

1. Ensure the pump has run for 5 minutes before verifying flow rate.
2. Ensure representative impactor is loaded with a support and collection filter and that it is fully assembled. *See Sampler Preparation on page 3.*
3. Align the bottom of the calibration adapter with the inlet plate of an assembled, loaded representative Disposable PPI and press down firmly until the calibration adapter's O-ring is engaged and creates an even seal.
4. Use flexible tubing to connect the exhaust of the Disposable PPI to the inlet of a sample pump.
5. Use flexible tubing to connect the inlet of the calibration adapter to the suction port (outlet) of a flowmeter.
6. Verify that flow rate is 2 L/min for the 2 L/min respirable and thoracic model PPIs, 4 L/min for the 4 L/min respirable model, or 8 L/min for the 8 L/min respirable model. Follow the instructions in the pump and flowmeter operating instructions.
7. When flow rate is verified, disconnect the tubing from the flowmeter and calibration adapter.
8. Grasp the Disposable PPI with one hand and the calibration adapter with the other hand. Firmly pull/twist to remove the calibration adapter from the Disposable PPI inlet plate.
9. Replace the representative Disposable PPI used to set the flow with a new loaded Disposable PPI for sample collection.



Using excessive pressure to seal the calibration adapter to the Disposable PPI may make the calibration adapter difficult to remove.




Sampling



SKC tests indicate that a particulate mass of up to 6.8 mg on the four impaction substrates would not affect PPI performance. This amount is equivalent to sampling for 6 hours at 4 L/min in environments where respirable mass concentration is 4.76 mg/m³ and equals 50% of total dust. However, labs have reported to SKC that they prefer no more than 2 mg on the filter for analytical reasons. Therefore, SKC recommends that you work with your lab to determine optimum sample times for your unique sampling conditions.

1. As per good industrial hygiene practice, replace representative sampler used for flow rate verification with a new, unused loaded sampler. See *Media and Sampler Preparation*.
2. Record sample start time on label.
3. Clip Disposable PPI onto a worker's collar or lapel in the breathing zone or in the area to be sampled.
4. Clip sample pump at the worker's waist or close to the Disposable PPI.
5. Use flexible tubing to attach the Disposable PPI exhaust to the inlet of the sample pump.
6. Turn on pump and record pertinent sample data.
7. After the desired sample time has elapsed, turn off the pump and unclip sampler from sampling location.
8. Apply Post-use label to inlet plate to seal the sampler.
9. Record sample stop time on label on the bottom of sampler.
10. Disconnect sampler from pump and reinstall protective end cap on the exhaust.
11. Reinstall flow rate verification train with representative Disposable PPI and verify flow rate. See *Flow Rate Verification*.



Sample Transport and Analysis

Package and transport samples and blanks to an accredited laboratory for gravimetric or chemical analysis.

References

Trakumas, S., Hall, P., *Personal Respirable Sampler Containing Four Impactors Arranged in Parallel, Abstracts of 23rd Annual AAAR Conference, Atlanta, GA, 2004*, p. 78

Trakumas, S., Salter, E., "Parallel Particle Impactor - Novel Size-selective Particle Sampler for Accurate Fractioning of Inhalable Particles," *Journal of Physics: Conference Series 151* (2009), 16 pp., 012060, [www.skinc.com/instructions/Parallel Particle Impactor Paper.pdf](http://www.skinc.com/instructions/Parallel%20Particle%20Impactor%20Paper.pdf)

Reference 2 is an author-created, non-copyrighted version of an article accepted for publication in the *Journal of Physics; Conference Series 151*. IOP Publishing Ltd. is not responsible for any errors or omissions in this version of the manuscript or any version derived from it. The definitive publisher authenticated version is available online. Go to <http://dx.doi.org>, enter doi: 10.1088/1742-6596/151/1/012060.

Trakumas, S., "High-flow Personal Respirator Dust Sampler for Increased Sensitivity," Poster 261, AIHce 2010, Denver, CO

Trakumas, S., "High-flow Personal Sampler to Monitor Exposure to Respirable Crystalline Silica at New Lower TLV," IOHA 2010 8th Conference Book of Abstracts, Rome, p. 59

Trakumas, S., Salter, E., "High-Flow Personal Sampler to Monitor Exposure to Respirable Crystalline Silica at New Lower TLV" PowerPoint Presentation

OSHA Final Rule on Respirable Crystalline Silica, www.osha.gov/silica/

MSHA Final Rule on Respirable Crystalline Silica, www.msha.gov/silica/

ISO 7708:1995 (2008), *Air Quality — Particle Size Fraction Definitions for Health-related Sampling*, www.iso.org, search on 7708

Stacey, P., Thorpe, A., and Echt, A., "Performance of High Flow Rate Personal Respirable Samplers When Challenged with Mineral Aerosols of Different Particle Size Distributions," *Ann. Occup. Hyg.*, 60, 2016, pp. 479-492, <http://annhyg.oxfordjournals.org/content/60/4/479.full.pdf>

Görner, P., Simon, X., Boivin, A., Bau, S., "Sampling Efficiency and Performance of Selected Thoracic Aerosol Samplers," *Annals of Work Exposure and Health*, 2017, Vol. 61, No. 7, 784-796

Ordering Information

User-loaded Disposable Plastic PPI Samplers , <i>select the PPI for the desired convention. Designed for one-time use.</i>	Cat. No.
User-loaded Disposable PPI Samplers contain four porous disc impaction substrates. <i>Require collection filter and support; see information below and select based on application</i>	
Respirable PPI (red) , 8 L/min, plastic	225-384
Respirable PPI (orange) , 4 L/min, plastic	225-387
Respirable PPI (gold) , 2 L/min, plastic	225-385
Thoracic PPI (blue) , 2 L/min, plastic	225-386
Recommended Collection Filters for User-loaded Disposable PPI , <i>required for sampling. Select a filter based on your application.</i>	
PVC Filters , 37 mm, 5.0- μ m pore size, pk/100	225-5-37
MCE Filters , 37 mm, 0.8- μ m pore size, pk/100	225-1939
Filter Supports , <i>required for User-loaded Disposable PPI sampling. Select either cellulose or stainless steel.</i>	
Support Pads , cellulose, 37 mm, pk/100	225-27
Support Pads , stainless steel, 37 mm, wide mesh, ea	225-26
Accessories	
Calibration Adapter for Disposable PPI	225-389
Forceps , stainless steel, non-serrated flat tips	225-8371

* Back pressure on PTFE filters can vary within the same lot.

SKC Limited Warranty and Return Policy

SKC products are subject to the SKC Limited Warranty and Return Policy, which provides SKC's sole liability and the buyer's exclusive remedy. To view the complete SKC Limited Warranty and Return Policy, go to skcinc.com/warranty.

skcinc.com

Form 38055 Rev 240719